

CERTIFICATE OF TYPE TESTING IN ACCORDANCE WITH EN 14175

CERTIFICATE & REPORT NO: INV/EN14175/619

DATE: 12th July 2013

Fume Cupboard Type:

SETLAB 150-01

1.5m wide bench type

External Dimensions:

Height = 2400mm
(1500mm + 900mm)

Width = 1500mm

Depth = 900mm

Internal Dimensions:

Height = 1200mm

Width = 1497mm

Depth = 760mm (wall-sash)

695mm (baffle-sash)

Baffle Gap Dimensions:

Top gap: 23mm

Bottom gap: 250mm

Side gap: 20mm

Fume Cupboard Manufacturer:

Set Laboratuvar Sist Ltd Sti

Mevlana Mah

Fatih Cad, No 24/1

Atasehir, Istanbul

TURKEY



Fume Cupboard Flow:

Volume flow rate: 0.384m³/s

Test opening width: 1420mm

Test opening height: 500mm

Face velocity: 0.502m/s

This is to certify that the fume cupboard described above has been type tested in accordance with Part 3, in compliance with the requirements of Part 2 and with reference to Part 1 of EN 14175 and resulted in performance characteristics given in the corresponding test report.

Tested and Certified by: Dr A F Bicen

Invent UK Ltd, 85 Southdown Road, Harpenden, Herts AL5 1PR

Tel: 01582 461666 :

A. F. Bicen